

RFP No: 26-005
BUYER: Kristy Jennings, Finance Accountant

Date: 06/09/2026
Email: Jennings-k@mssu.edu

ADDENDUM #1

Please submit this signed addendum with your bid response.

Please Note: See below for clarification and additional information:

1. Why must the Wall and Table Bucky grids be removable? We have systems with both – most systems built today commonly have fixed grids in the wall-stand and table bucky and a software grid is used outside of the buckys. This choice limits your selection of models available.
 - a. This is so we are able to visualize and show grid errors with manual changes for student involvement and visualization.
2. Would you accept 2-17"x17" detectors in the configurations? Some of our systems only have 17"x17" detectors.
 - a. We want the versatility of honing technologist skills to both a 14X17 and 17X17 for clinical application. Positive Beam Limitation would be acceptable for the with a 17X17.
3. Are these system's tubes – ceiling supported or floor supported – or 1 of each? I believe we have previously quoted a floor mounted system for MSSU's use-but that's not outlined in the RFP. Please clarify.
 - a. We are looking for 1 ceiling support and 1 floor support x-ray system to have minimal construction completed on the existing rooms.
4. Are both these rooms both 3 phase electrical or is one single phase?
 - a. There are both 3 phase electrical systems.
5. What generator size is needed for these 2 rooms? Saw nothing in the RFP about generator capabilities.
 - a. Generator size dependent on Vendor recommendation.
6. Do you have room dimensions and ceiling heights for these 2 rooms that could be provided? Or a CAD File?
 - a. We do not, room dimensions can be obtained during the mandatory pre-bid meeting on 6/15.

7. Could I make an appointment to come onsite next week to submit the trade-in/deinstallation information for your current equipment?
 - a. Any trade-in/Deinstallation information should be included in your bid package.

8. UPS System Included – is this a console UPS or a full system UPS?
 - a. Console UPS at minimum

9. Full installation and training completed by September 15, 2026 (or earlier) – is this a firm timeline? I interpret this to include site construction, equipment delivery, service installation, and clinical trainings. Is that correct?
 - a. Yes, if at all possible. We recognize there may be other time constraints, but prefer the September deadline.

10. Evaluates 50 year cost, including consumables and high-value replacements – in the evaluation grid, its states 50 year cost in column 1 but in a different column states 5 year. Was 50 year a typo?
 - a. Yes, 50 year was a typo. We meant 5 year.

This Addendum is hereby considered to be a part of the original bid specifications and needs to be signed and returned with your bid in order to be considered.

Signature

Date